



IGC AGENCY cc T/A

IGC SCALES

WEIGHING EQUIPMENT - INDUSTRIAL & RETAIL SALES & SERVICE

Est. 1972



LTF 0099

55 BOK STREET / ☒ 998
POLOKWANE 0700
E-mail: mwigc@mweb.co.za

☎ 015 295 4164 / 205 5537
FAX: 015 295 4164

VERIFICATION CERTIFICATE TEST REPORT



Customer: <u>VUKA-KHULANI TIMBER INDUSTRIES</u> <u>OLD STATION ROAD</u>	Address: <u>P.O Box 613</u> <u>Durbanville 6035</u>	Certificate No: <u>JK#31286</u>
---	---	------------------------------------

SCALE DETAILS			
Make : <u>AVER1</u>	Capacity : <u>5000kg</u>	New :	<input type="checkbox"/>
Model : <u>L115</u>	Division Size : <u>10kg</u>	Repair :	<input type="checkbox"/>
Serial No : <u>H99290760</u>	S.A. Approval No : <u>1207</u>	In Trade use :	<input checked="" type="checkbox"/>

PERMISSABLE TOLERANCES			
<u>0kg</u>	TO	<u>5000kg</u>	<u>0kg</u>
<u>5000kg</u>	TO	<u>2000kg</u>	<u>20kg</u>
<u>2000kg</u>	TO	<u>3000kg</u>	<u>30kg</u>
	TO		
	TO		
	TO		
	TO		

ADDITIONAL TESTS	
Price Computing :	<u>N/A</u>
Tare :	<u>C</u>
Environmental Conditions :	<u>C</u>
Limits of Indication :	<u>C</u>
Lateral Tests :	<u>N/A</u>
Printer :	<u>N/A</u>
Discrimination :	<u>C</u>
Repeatability :	<u>C</u>

ACCURACY TEST			
MASS	↑	ERRORS	↓
<u>0kg</u>	<u>C</u>		<u>C</u>
<u>200kg</u>	<u>C</u>		<u>C</u>
<u>500kg</u>	<u>C</u>		<u>C</u>
<u>1000kg</u>	<u>C</u>		<u>C</u>
<u>1500kg</u>	<u>C</u>		<u>C</u>
<u>2000kg</u>	<u>C</u>		<u>C</u>
<u>2500kg</u>	<u>C</u>		<u>C</u>
<u>3000kg</u>	<u>C</u>		<u>C</u>

Observations made during the preliminary Examination: Scale correct.

C = CORRECT	R = REJECT	N/A = NOT APPLICABLE
--------------------	-------------------	-----------------------------

STANDARDS

The mass pieces used for this verification are traceable to National standards through an approved SANAS laboratory

Calibration Cert No: 65/09/886
 Date of Calibration: MAY 09
 ID of Mass: C+S MASS
 Own:
 Loan:

END	MIDDLE	END
ECCENTRIC TEST		
<u>10000kg</u>	_____	<u>10000kg</u>
<u>10000kg</u>	_____	<u>10000kg</u>

VERIFICATION OFFICER

Name : S. de Beer

Seal no : IGC 3

Verification Procedure : IGC/VP/01

Signature : [Signature]

This instrument was tested and found to comply in all respects with the requirements of The Trade Metrology Act, 1973 (Act No. 77 of 1973) and may be used for a prescribed Purpose as included in the Act.

Date 11/6/09 Expiry date of Certificate 11/6/11